

Section VII. Technical Specifications

Technical Specifications

| Item | Specification | Statement of Compliance |
|-------------|----------------------|--|
| | | <i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i> |

| ITEM NO. | ITEMS | SPECIFICATIONS | UNIT | STATEMENT OF COMPLIANCE |
|-----------------|-----------------------------|---|-------------|--------------------------------|
| 1 | Infusion Administration Set | with air vent, without needle, with micron filter, clear tubing 178-244cm, Adult | set | |
| 2 | Infusion Administration Set | with air vent, without needle, with micron filter, clear tubing 178-244cm, Pedia | set | |
| 3 | Blood Transfusion set | DEHP free, with clear, needleless injection site, 145-245cm tubing luer lock with G.18 or G.19 needle | set | |
| 4 | Precision drip chamber | Calibrated burette, 60 drops/ml. (can regulate to the slowest rate of 2 cc/hour) 120-150ml | set | |

| | | | | |
|---|----------------------|---------------------------------------|-----|--|
| 5 | Intravenous Catheter | g. 18 with wings, with injection port | set | |
| 6 | Intravenous Catheter | g. 20 with wings, with injection port | set | |
| 7 | Intravenous Catheter | g. 22 with wings, with injection port | set | |
| 8 | Intravenous Catheter | g. 24 with wings, with injection port | set | |
| 9 | Intravenous Catheter | g. 26 with wings, with injection port | set | |

TERMS & CONDITIONS:

1. The quantities specified are estimated quantities only and may be decreased based on the actual need of the hospital and availability of fund. It is understood that EAMC is not bound to purchase all of the quantities/items called for in the Invitation to Bid (IB).
2. Deliveries shall have at least eighteen (18) months expiration date/shelf-life. For items with shorter expiration date, the Supplier shall submit an assurance letter prior to deliver for approval of the Head of the Procuring Entity (HOPE) or his authorized representative.

| |
|---|
| <i>Conforme:</i> |
| <i>Company/Bidder's Name:</i> |
| <i>Name and Signature of Authorized Representative:</i> |
| <i>Date:</i> |