

Section VII. Technical Specifications

Technical Specifications

Item	Specification	Statement of Compliance
		<i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>

ITEM NO.	ITEMS	SPECIFICATIONS	UNIT	STATEMENT OF COMPLIANCE
Cotton				
1	Cotton	absorbent, 5 pieces per pack, sterile	pack	
Heart Station, Neuroscience, and Rehabilitation Medicine use				
2	ECG Electrodes	adult, large size, good adhesive, good conductivity	piece	
3	ECG Electrodes	neonate/pedia size, good adhesive, good conductivity	piece	
4	ECG Patches	Adult, 30 pieces/pouches	pouch	

5	EEG conductive paste	400g	jar	
6	Gel	ultrasound transmission	gallon	
7	Mineral oil	for paraffin wax mixture, 1 liter	gallon	
8	Skin preparation gel for Neuro	114-135g	tubes	
9	Paper	thermal, size: 50mm x 30mm for Kenz ECG machine	roll	
10	Paper	thermal, size: 90mm x 80mm fanfold, for Zoll defibrillator	roll	
11	Paper	thermal, black and white, high quality, 110mm/20m, type I	roll	
12	Paper	thermal, size: 60mmx15m, for AR600 Cardioline ECG machine	roll	
13	Paper	thermal, size: 63mmx30m, for Nihon Kohden ECG machine	roll	
14	Paper	for otoacoustic emission test, 57mmx30m, 1/2 core	roll	
15	Wax paper	for physical therapy procedures	roll	
16	Wax pellets	for physical therapy procedures	kilo	
IV Splint (Lot Bidding)				
17	IV Splint	padded, with strap, disposable, adult, large	piece	
18	IV Splint	padded, with strap, disposable, pedia	piece	
Ostomy Supplies				
19	Ostomy Starter Kit	Ostomy Starter Kit 45mm, 1 Wafer Barrier, 1 Drainable Pouch Bag, 1 tail clip, 1 measuring guide	kit	
20	Ostomy Starter Kit	Ostomy Starter Kit 57mm, 1 Wafer Barrier, 1 Drainable Pouch Bag, 1 tail clip, 1 measuring guide	kit	
21	Ostomy Starter Kit	Ostomy Starter Kit 70mm, 1 Wafer Barrier, 1 Drainable Pouch Bag, 1 tail clip, 1 measuring guide	kit	
Sterilization Indicators and Controls for Low Temperature Steam Formaldehyde Sterilizer				
22	Solution bag	sterilant solution bag, plastic disposable bags of 2.7 litres capacity with only 2% of formaldehyde concentration, boxes of 3 bags, for Matachana	box	
23	Indicators	self-adhesive indicators spots (1000UU) Chemical Type 1 process indicator, for steam sterilization processes formaldehyde (LTSF). Chemical indicator dots to use as external verification in order to check sterilization processes. The colour is	box	

		changing from golden yellow changes to strong purple or black. Expiration date: 5 years, for Matachana		
24	Indicators	LTSF tape for formaldehyde sterilization, 24's/ box, Special tape for LTSF sterilization. With indicator of class 1 on its surface that gets a colour change in presence of formaldehyde. Packed in individual packs. boxes of 24 units	box	
25	Strips	process challenge device (Helix), Chemical Type 4 validation test kit + 200 indicating strips for formaldehyde process check: 1. - PCD: Process challenge device in helix format. 2. - 200 indicating strips with a specific chemical reaction for formaldehyde. Self- adhesive to keep them on the registrar sheet. Developed and manufactured according to EN ISO 11140- 1: 2015. It is validated for penetration performance testing low temperature steam formaldehyde (LTSF) sterilization processes. Chemical strips are designed to change colour from golden yellow to strong purple or black. It is easy to use and interpret. Diagnostic capabilities include air extraction and penetration performance. Expiration Date: 2 years, for Matachana	box	
26	Printer paper	for LTSF, Folded printer paper, for Matachana	piece	
Surgical Drains and Reservoir				
27	Drain	silicon flat with reservoir of 100-150ml	piece	
28	Penrose drain	surgical, rubber, size: 1/2	piece	
29	Penrose drain	surgical, rubber, size: 5/8	piece	
30	T Tube drain	thin latex, fr 22	piece	
31	T Tube drain	thin latex, fr 20	piece	
32	T Tube drain	thin latex, fr 18	piece	
33	T Tube drain	thin latex, fr 16	piece	
Surgical Ligating Clips				
34	Hemolok ligating clips (medium)	With bosses, integrated teeth, locking mechanism, non absorbable large (green)	cartridge	
35	Hemolok ligating clips (large)	With bosses, integrated teeth, locking mechanism, non absorbable large (violet)	cartridge	
36	Hemolok ligating clips (extra large)	With bosses, integrated teeth, locking mechanism, non absorbable large (gold)	cartridge	
37	Titanium	Chevron shaped, heart shaped wires, with	cartridge	

	ligating clips, medium to large	internal grooves, 6 clips/cartridge, 20 cartridges/box		
38	Titanium ligating clips, large	Chevron shaped, heart shaped wires, with internal grooves, 6 clips/cartridge, 20 cartridges/box	cartridge	
Topical Hemostatic Items				
39	Bone wax	beeswax, 2,5 grams, sterile, box of 24	box	
40	Hemostat	absorbable gelatin sponge, 8cmx5cmx1cm	piece	
41	Hemostat	oxidized regenerated cellulose, Size 3" x 4	piece	
42	Hemostat	oxidized regenerated cellulose, Size 2" x 4	piece	

TERMS & CONDITIONS:

1. The quantities specified are the estimated quantities only and may be decreased based on the actual need of the hospital and availability of fund. It is understood that EAMC is not bound to purchase all of the quantities/items called for in Invitation to Bid (ITB).
2. Deliveries shall have at least eighteen (18) months expiration date/shelf-life. For items with shorter expiration date, the Supplier shall submit an assurance letter prior to deliver for approval of the Head of the Procuring Entity (HOPE) or his authorized representative.

<i>Conforme:</i>
<i>Company/Bidder's Name:</i>
<i>Name and Signature of Authorized Representative:</i>
<i>Date:</i>